

Table of Contents

1.0 Purpose/Objective	2
2.0 General Scope	2
3.0 Definitions	2
4. DC/Low Frequency Transfer Standards	3
4.1 Fluke 87IV Multimeter	3
5.0 RF Transfer Standards up to 40 GHz	5
5.1 Spanawave/Gigatronics 8541C RF Power Meter with Spanawave/ Gigatronics 80324A Power Sensor	5
5.2 Micronetics NS346KA RF Noise Power Source	6
5.3 Anritsu SC7777 Open Short Load Standards (OSL K-Connector)	6
6.0 V-Band Transfer Standards 50-75 GHz	7
6.1 Keysight N1913A 80 GHz RF Power Meter with Keysight V8486A 50-75 GHz Power Sensor	7
6.2 OML WR15 Waveguide Offset Standards	9
7.0 E-Band Transfer Standards 60-88 GHz	10
7.1 Keysight N1913A 80 GHz RF Power Meter with Keysight E8486A-100 60-90 GHz Power Sensor	10
7.2 OML WR12 Waveguide Offset Standards	12

Roos Instruments, Inc.
2285 Martin Avenue
Santa Clara, California 95050
United States of America
www.roos.com
408-748-8589

1.0 Purpose/Objective

This document outlines the measurement capabilities and an estimation of measurement uncertainties and/or errors for calibration with the Roos Instruments, Inc. manufacturing and field service operations.

2.0 General Scope

The electronic and physical measurement limitations and characteristics of the Roos Instruments calibration transfer standards are used in all manufacturing and field service operations. Values shown represent best measurement capabilities.

3.0 Definitions

3.1 Calibration: A set of procedures that when executed establish under specific conditions the relationship between the values indicated by a measuring instrument and the corresponding know characteristics of a particular transfer standard.

3.2 Transfer Standard: A physical or electrical device that has its performance characteristics measured, documented and compared to a know reference at Primary calibration laboratory such as US National Institute of Science and Technology (NIST).

3.3. Measurement Uncertainty: Calculated estimation of maximum errors possible in measurement accuracy at a specific point or function for a given measuring instrument.

For more information:

roos.com/support

4. DC/Low Frequency Transfer Standards

4.1 Fluke 87IV Multimeter

Measurements	Uncertainty Label	Uncertainty Value
Voltage DC	Maximum voltage	1000 V
	Accuracy	$\pm(0.05\% + 1)$
	Maximum resolution	10 μ V
Voltage AC	Maximum voltage	1000 V
	Accuracy	$\pm(0.7\% + 2)$ True RMS
	AC bandwidth	20 kHz with low pass filter; 3 db @ 1 kHz
	Maximum resolution	0.1 mV
Current DC	Maximum amps	10 A (20 A for 30 seconds maximum)
	Amps accuracy	$\pm(0.2\% + 2)$
	Maximum resolution	0.01 μ A
Current AC	Maximum amps	10 A (20 A for 30 seconds maximum)
	Amps accuracy	$\pm(1.0\% + 2)$ True RMS
	Maximum resolution	0.1 μ A
Resistance	Maximum resistance	50 M Ω
	Accuracy	$\pm(0.2\% + 1)$
	Maximum resolution	0.1 Ω
Capacitance	Maximum capacitance	9,999 μ F
	accuracy	$\pm(1\% + 2)$
	Maximum resolution	0.01 nF
Frequency	Maximum frequency	200 kHz
	Accuracy	$\pm(0.005\% + 1)$
	Maximum resolution	0.01 Hz
Duty cycle	Maximum duty cycle	99.9%
	Accuracy	$\pm(0.2\%$ per khz $+ 0.1\%)$
	Maximum resolution	0.1%
Temperature measurement	-200.0°C –1090°C -328.0°F –1994.0°F excluding probe	
80 BK temperature probe	-40.0°C –260°C -40.0°F –500°F, 2.2°C or 2% whichever is greater	

Measurements	Uncertainty Label	Uncertainty Value
Conductance	Maximum conductance	60.00 nS
	Accuracy	$\pm(1.0\% + 10)$
	Maximum resolution	0.01 nS
Diode	Range	3 V
	Resolution	1 mV
	Accuracy	$\pm(2\% + 1)$
Duty cycle range	Accuracy	Within $\pm(0.2\%$ per kHz + 0.1%)

Environmental Specifications	
Operating temperature	-°C to + 55°C
Storage temperature	-°C to + 60°C
Humidity (without condensation)	0% – 90% (0°C – 35°C) 0% – 70% (35°C – 55°C)
Operating Altitude	2000 m

Safety Specifications	
Overvoltage category	EN 61010 to 1000 V CAT III, 600V CAT IV
Agency approvals	UL, CSA, TÜV, VDE listed

5.0 RF Transfer Standards up to 40 GHz

5.1 Spanawave/Gigatronics 8541C RF Power Meter with Spanawave/Gigatronics 80324A Power Sensor

The accuracy calculation table lists the significant uncertainties of an absolute power measurement. The accuracy of the 8541C combined with the 80324A sensor is shown at +20 dBm, 0 dBm, and -30 dBm; Frequency = 1 GHz; Source Match = 1.5:1.

Measurements	Uncertainty at 1 GHz, 1.5:1 Source Match		
	@ 20 dBm	@ 0 dBm	@ -30 dBm
8541C with 80324A			
Instrumentation Uncertainty	±5.2%	±0%	±0.925%
Sensor Power Linearity (>8 GHz)	±0%	±0%	±0%
Calibrator Uncertainty	±1.2%	±1.2%	±1.2%
Calibrator/Sensor Mismatch	±0.28%	±0.28%	±0.28%
Calibration Factor Uncertainty	±1.04%	±1.04%	±1.04%
Zero Error	±0.00000005%	±0.00000005%	±0.005%
Noise	±0.00000005%	±0.00000005%	±0.005%
Mismatch (Sensor/Source)	±2.25%	±2.25%	±2.25%
% Total Uncertainty	±9.97%	±4.77%	±5.71%
dB Total Uncertainty	±0.41 dB	±0.20 dB	±0.24 dB

5.2 Micronetics NS346KA RF Noise Power Source

Measurements	Uncertainty
RF Frequency	100 MHz to 40000 MHz
Noise Output	10 - 17 dB ENR
Noise Spectral Density	-160.5 dBm/Hz
VSWR	1.25:1 Max (5 - 12 GHz)
	1.30:1 Max (12 - 18 GHz)
	1.40:1 Max (18 - 26.5 GHz)
	1.50:1 Max (26.5 - 40 GHz)

5.3 Anritsu SC7777 Open Short Load Standards (OSL K-Connector)

Measurements	Uncertainty Label	Uncertainty Value
Reference Plane Pin Depth	FEMALE	0.207 ±0.003 inches (5.2578 ±0.0762 mm)
	MALE	-0.207 ±0.003 inches (-5.2578 ±0.0762 mm)

6.0 V-Band Transfer Standards 50-75 GHz

6.1 Keysight N1913A 80 GHz RF Power Meter with Keysight V8486A 50-75 GHz Power Sensor

The accuracy calculation table lists the significant uncertainties of an absolute power measurement. The accuracy of the N1913A combined with the V8486A sensor is shown at +20 dBm, 0 dBm, and -26 dBm; Frequency = 65 GHz; VSWR = 1.5:1.

Calculator Source Link: <https://community.keysight.com/servlet/JiveServlet/download/71731-1-6558/Absolute+Power+Sensor+Uncertainty.xlsx>

Revision 10

Date updated: 7 Sep 2017

© Keysight Technologies, Inc. 2012, 2014 and 2015.

The data used in this calculator is based on the specifications in the EPM, EPM and P-Series datasheets.

The calculations are based on the ISO Guide to the Expression of Uncertainty in Measurement, often referred to as the GUM.

For more info about the GUM, and models used in this calculator, please refer to Agilent Application Note 1449-3,

"Fundamentals of RF and Microwave Power Measurements (Part 3)," literature number 5988-9215EN.

65 GHz @ +20 dBm

Keysight Average Power Sensor Uncertainty Calculator		Instructions: Fill in all the blue-colored fields and the "Sensor Match" box; observe results in lower right.			
Input	Value	Match Specification Style [†]	Distribution of ρ [†]		
Device Under Test match (RL or SWR or ρ)*	1.22 VSWR	Maximum	C: Rayleigh		
Sensor Model	E8486A-100	Maximum	C: Rayleigh		
User-Entered Sensor Match*	1.50 VSWR	Maximum	C: Rayleigh		
Sensor Match at Measurement Frequency: Data Source is <input type="radio"/> From Sensor Model (via Specifications vs. Frequency) <input checked="" type="radio"/> User-Entered; use the line above this box		[†] These two columns set the form of the specification and the PDF of the match. Hover below for details.			
Power Meter Model	EPM				
Frequency	80 GHz				
Power	20 dBm		0.1 W		
Number of Readings Averaged	1				
Average mode	Normal				
Source of Uncertainty	Symbol	Value \pm	Probability Distribution	Divisor	Result
Mismatch Gain Between Generator and Sensor	M_u	$ \Gamma_{max} _g = 0.099$ $ \Gamma_{max} _s = 0.200$	C: Rayleigh C: Rayleigh	4.179	0.474%
Mismatch Gain Between Calibration Source and Sensor	M_{uc}	$ \Gamma _c = 0.029$ $ \Gamma_{max} _{sc} = 0.032$	A: Uniform inside Circle C: Rayleigh	2.431	0.038%
Power Meter Instrumentation Error	P_m	0.50%	Gaussian	2.000	0.250%
Power Meter Instrumentation Error During Calibration	P_{mc}	0.50%	Gaussian	2.000	0.250%
Power Meter Calibrator Output Power	P_{cal}	0.50%	Gaussian	2.000	0.250%
Zero Drift	D	4.000E-08	Gaussian	2.000	0.000%
Power Sensor Calibration Factor Uncertainties	K_b	5.300%	Gaussian	2.000	2.650%
Power Sensor Linearity	P_l	2.000%	Gaussian	2.000	1.000%
Zero Set	Z_s	2.000E-07	Gaussian	2.000	0.000%
Sensor Noise	N	4.500E-07	Gaussian	2.000	0.001%
				Combined Uncertainty-RSSed =	2.90%
				K =	2.00
				Expanded Uncertainty =	5.81%
				Upper Limit Uncertainty =	+ 0.245dB
				Lower Limit Uncertainty =	- 0.260dB

* This term can be entered as the return loss (RL) in dB (Sxx), or VSWR, or as a reflection coefficient. e.g. RL = -15 dB is equivalent to VSWR = 1.43 is equivalent to $\rho = 0.178$ (reflection coefficient)
Enter 1 for VSWR = 1, a perfect match, or 0 for RL = 0 dB, a lossless reflection

65 GHz @ 0 dBm

Keysight Average Power Sensor Uncertainty Calculator		Instructions: Fill in all the blue-colored fields and the "Sensor Match" box; observe results in lower right.			
Input	Value	Match Specification Style[†]	Distribution of ρ[†]		
Device Under Test match (RL or SWR or ρ)*	1.22 VSWR	Maximum	C: Rayleigh		
Sensor Model	E8486A-100	Maximum	C: Rayleigh		
User-Entered Sensor Match*	1.50 VSWR	Maximum	C: Rayleigh		
<div style="border: 1px solid black; padding: 5px;"> Sensor Match at Measurement Frequency: Data Source is <input type="radio"/> From Sensor Model (via Specifications vs. Frequency) <input checked="" type="radio"/> User-Entered; use the line above this box </div>		[†] These two columns set the form of the specification and the PDF of the match. Hover below for details.			
Power Meter Model	EPM				
Frequency	80 GHz				
Power	0 dBm			0.001 W	
Number of Readings Averaged	1				
Average mode	Normal				
Source of Uncertainty	Symbol	Value \pm	Probability Distribution	Divisor	Result
Mismatch Gain Between Generator and Sensor	M_u	$ \Gamma_{max} _g = 0.099$ $ \Gamma_{max} _s = 0.200$	C: Rayleigh C: Rayleigh	4.179	0.474%
Mismatch Gain Between Calibration Source and Sensor	M_{uc}	$ \Gamma _c = 0.029$ $ \Gamma_{max} _{sc} = 0.032$	A: Uniform inside Circle C: Rayleigh	2.431	0.038%
Power Meter Instrumentation Error	P_m	0.50%	Gaussian	2.000	0.250%
Power Meter Instrumentation Error During Calibration	P_{mc}	0.50%	Gaussian	2.000	0.250%
Power Meter Calibrator Output Power	P_{cal}	0.50%	Gaussian	2.000	0.250%
Zero Drift	D	4.000E-08	Gaussian	2.000	0.002%
Power Sensor Calibration Factor Uncertainties	K_b	5.300%	Gaussian	2.000	2.650%
Power Sensor Linearity	P_l	1.000%	Gaussian	2.000	0.500%
Zero Set	Z_s	2.000E-07	Gaussian	2.000	0.010%
Sensor Noise	N	4.500E-07	Gaussian	2.000	0.124%
				Combined Uncertainty-RSSed =	2.78%
				K =	2.00
* This term can be entered as the return loss (RL) in dB (Sxx), or VSWR, or as a reflection coefficient. e.g. RL = -15 dB is equivalent to VSWR = 1.43 is equivalent to ρ = 0.178 (reflection coefficient) Enter 1 for VSWR = 1, a perfect match, or 0 for RL = 0 dB, a lossless reflection				Expanded Uncertainty =	5.55%
				Upper Limit Uncertainty =	+ 0.235dB
				Lower Limit Uncertainty =	- 0.248dB

65 GHz @ -26 dBm

Keysight Average Power Sensor Uncertainty Calculator		Instructions: Fill in all the blue-colored fields and the "Sensor Match" box; observe results in lower right.			
Input	Value	Match Specification Style[†]	Distribution of ρ[†]		
Device Under Test match (RL or SWR or ρ)*	1.22 VSWR	Maximum	C: Rayleigh		
Sensor Model	E8486A-100	Maximum	C: Rayleigh		
User-Entered Sensor Match*	1.50 VSWR	Maximum	C: Rayleigh		
Sensor Match at Measurement Frequency: Data Source is <input type="radio"/> From Sensor Model (via Specifications vs. Frequency) <input checked="" type="radio"/> User-Entered; use the line above this box		[†] These two columns set the form of the specification and the PDF of the match. Hover below for details.			
Power Meter Model	EPM				
Frequency	80 GHz				
Power	-26 dBm			2.51189E-06 W	
Number of Readings Averaged	1				
Average mode	Normal				
Source of Uncertainty	Symbol	Value \pm	Probability Distribution	Divisor	Result
Mismatch Gain Between Generator and Sensor	M_u	$ \Gamma_{max} _g = 0.099$ $ \Gamma_{max} _s = 0.200$	C: Rayleigh C: Rayleigh	4.179	0.474%
Mismatch Gain Between Calibration Source and Sensor	M_{uc}	$ \Gamma _c = 0.029$ $ \Gamma_{max} _{sc} = 0.032$	A: Uniform inside Circle C: Rayleigh	2.431	0.038%
Power Meter Instrumentation Error	P_m	0.50%	Gaussian	2.000	0.250%
Power Meter Instrumentation Error During Calibration	P_{mc}	0.50%	Gaussian	2.000	0.250%
Power Meter Calibrator Output Power	P_{cal}	0.50%	Gaussian	2.000	0.250%
Zero Drift	D	4.000E-08	Gaussian	2.000	0.796%
Power Sensor Calibration Factor Uncertainties	K_b	5.300%	Gaussian	2.000	2.650%
Power Sensor Linearity	P_l	1.000%	Gaussian	2.000	0.500%
Zero Set	Z_s	2.000E-07	Gaussian	2.000	3.981%
Sensor Noise	N	4.500E-07	Gaussian	2.000	49.266%
				Combined Uncertainty-RSSed =	49.51%
				K =	2.00
* This term can be entered as the return loss (RL) in dB (Sxx), or VSWR, or as a reflection coefficient. e.g. RL = -15 dB is equivalent to VSWR = 1.43 is equivalent to $\rho = 0.178$ (reflection coefficient) Enter 1 for VSWR = 1, a perfect match, or 0 for RL = 0 dB, a lossless reflection			Expanded Uncertainty = 99.02% Upper Limit Uncertainty = + 2.989dB Lower Limit Uncertainty = - 20.092dB		

6.2 OML WR15 Waveguide Offset Standards

+25 C	MIN	TYP	MAX
System Operating Frequency (WR-15) ¹	50 GHz	--	75 GHz
Length of Shim (1/4 wavelength)	.0635 +/- .0002 in		
Return Loss of Fixed Load, Adjustable Load	--	>35 dB	--
Damage Level	--	+13 dBm	--
Operating Temperature Range	+20 °C	+25 °C	+30 °C

¹Test Port Flange Configuration is compatible with MIL-DTL-3922/67D (UG387/U-M)

7.0 E-Band Transfer Standards 60-88 GHz

7.1 Keysight N1913A 80 GHz RF Power Meter with Keysight E8486A-100 60-90 GHz Power Sensor

The accuracy calculation table lists the significant uncertainties of an absolute power measurement. The accuracy of the N1913A combined with the E8486A sensor is shown at +20 dBm, 0 dBm, and -26 dBm; Frequency = 80 GHz; VSWR = 1.5:1.

Calculator Source Link: <https://community.keysight.com/servlet/JiveServlet/download/71731-1-6558/Absolute+Power+Sensor+Uncertainty.xlsx>

Revision 10

Date updated: 7 Sep 2017

© Keysight Technologies, Inc. 2012, 2014 and 2015.

The data used in this calculator is based on the specifications in the EPM, EPM and P-Series datasheets.

The calculations are based on the ISO Guide to the Expression of Uncertainty in Measurement, often referred to as the GUM.

For more info about the GUM, and models used in this calculator, please refer to Agilent Application Note 1449-3,

"Fundamentals of RF and Microwave Power Measurements (Part 3)," literature number 5988-9215EN.

+20 dBm @ 80 GHz

Keysight Average Power Sensor Uncertainty Calculator		Instructions: Fill in all the blue-colored fields and the "Sensor Match" box; observe results in lower right.			
Input	Value	Match Specification Style [†]	Distribution of ρ [†]		
Device Under Test match (RL or SWR or ρ)*	1.22 VSWR	Maximum	C: Rayleigh		
Sensor Model	E8486A-100	Maximum	C: Rayleigh		
User-Entered Sensor Match*	1.50 VSWR	Maximum	C: Rayleigh		
<input type="radio"/> Sensor Match at Measurement Frequency: Data Source is <input type="radio"/> From Sensor Model (via Specifications vs. Frequency) <input checked="" type="radio"/> User-Entered; use the line above this box		[†] These two columns set the form of the specification and the PDF of the match. Hover below for details.			
Power Meter Model	EPM				
Frequency	80 GHz				
Power	20 dBm		0.1 W		
Number of Readings Averaged	1				
Average mode	Normal				
Source of Uncertainty	Symbol	Value \pm	Probability Distribution	Divisor	Result
Mismatch Gain Between Generator and Sensor	M_u	$ \Gamma_{max} _g = 0.099$ $ \Gamma_{max} _s = 0.200$	C: Rayleigh C: Rayleigh	4.179	0.474%
Mismatch Gain Between Calibration Source and Sensor	M_{uc}	$ \Gamma _c = 0.029$ $ \Gamma_{max} _{sc} = 0.032$	A: Uniform inside Circle C: Rayleigh	2.431	0.038%
Power Meter Instrumentation Error	P_m	0.50%	Gaussian	2.000	0.250%
Power Meter Instrumentation Error During Calibration	P_{mc}	0.50%	Gaussian	2.000	0.250%
Power Meter Calibrator Output Power	P_{cal}	0.50%	Gaussian	2.000	0.250%
Zero Drift	D	4.000E-08	Gaussian	2.000	0.000%
Power Sensor Calibration Factor Uncertainties	K_b	5.300%	Gaussian	2.000	2.650%
Power Sensor Linearity	P_l	2.000%	Gaussian	2.000	1.000%
Zero Set	Z_s	2.000E-07	Gaussian	2.000	0.000%
Sensor Noise	N	4.500E-07	Gaussian	2.000	0.001%
Combined Uncertainty-RSSed =					2.90%
K =					2.00
* This term can be entered as the return loss (RL) in dB (Sxx), or VSWR, or as a reflection coefficient. e.g. RL = -15 dB is equivalent to VSWR = 1.43 is equivalent to $\rho = 0.178$ (reflection coefficient) Enter 1 for VSWR = 1, a perfect match, or 0 for RL = 0 dB, a lossless reflection					Expanded Uncertainty = 5.81%
					Upper Limit Uncertainty = +0.245dB
					Lower Limit Uncertainty = -0.260dB

0 dBm @ 80 GHz

Keysight Average Power Sensor Uncertainty Calculator		Instructions: Fill in all the blue-colored fields and the "Sensor Match" box; observe results in lower right.			
Input	Value	Match Specification Style[†]	Distribution of ρ[†]		
Device Under Test match (RL or SWR or ρ)*	1.22 VSWR	Maximum	C: Rayleigh		
Sensor Model	E8486A-100	Maximum	C: Rayleigh		
User-Entered Sensor Match*	1.50 VSWR	Maximum	C: Rayleigh		
<div style="border: 1px solid black; padding: 5px;"> Sensor Match at Measurement Frequency: Data Source is <input type="radio"/> From Sensor Model (via Specifications vs. Frequency) <input checked="" type="radio"/> User-Entered; use the line above this box </div>		[†] These two columns set the form of the specification and the PDF of the match. Hover below for details.			
Power Meter Model	EPM				
Frequency	80 GHz				
Power	0 dBm			0.001 W	
Number of Readings Averaged	1				
Average mode	Normal				
Source of Uncertainty	Symbol	Value \pm	Probability Distribution	Divisor	Result
Mismatch Gain Between Generator and Sensor	M_u	$ \Gamma_{max} _g = 0.099$ $ \Gamma_{max} _s = 0.200$	C: Rayleigh C: Rayleigh	4.179	0.474%
Mismatch Gain Between Calibration Source and Sensor	M_{uc}	$ \Gamma _c = 0.029$ $ \Gamma_{max} _{sc} = 0.032$	A: Uniform inside Circle C: Rayleigh	2.431	0.038%
Power Meter Instrumentation Error	P_m	0.50%	Gaussian	2.000	0.250%
Power Meter Instrumentation Error During Calibration	P_{mc}	0.50%	Gaussian	2.000	0.250%
Power Meter Calibrator Output Power	P_{cal}	0.50%	Gaussian	2.000	0.250%
Zero Drift	D	4.000E-08	Gaussian	2.000	0.002%
Power Sensor Calibration Factor Uncertainties	K_b	5.300%	Gaussian	2.000	2.650%
Power Sensor Linearity	P_l	1.000%	Gaussian	2.000	0.500%
Zero Set	Z_s	2.000E-07	Gaussian	2.000	0.010%
Sensor Noise	N	4.500E-07	Gaussian	2.000	0.124%
				Combined Uncertainty-RSSed =	2.78%
				K =	2.00
				Expanded Uncertainty =	5.55%
* This term can be entered as the return loss (RL) in dB (Sxx), or VSWR, or as a reflection coefficient. e.g. RL = -15 dB is equivalent to VSWR = 1.43 is equivalent to $\rho = 0.178$ (reflection coefficient)				Upper Limit Uncertainty =	+ 0.235dB
Enter 1 for VSWR = 1, a perfect match, or 0 for RL = 0 dB, a lossless reflection				Lower Limit Uncertainty =	- 0.248dB

-26 dBm @ 80 GHz

Keysight Average Power Sensor Uncertainty Calculator

Input		Value	Match Specification Style[†]		Distribution of ρ[†]
Device Under Test match (RL or SWR or ρ)*		1.22 VSWR	Maximum		C: Rayleigh
Sensor Model		E8486A-100	Maximum		C: Rayleigh
User-Entered Sensor Match*		1.50 VSWR	Maximum		C: Rayleigh
Sensor Match at Measurement Frequency: Data Source is <input type="radio"/> From Sensor Model (via Specifications vs. Frequency) <input checked="" type="radio"/> User-Entered; use the line above this box			[†] These two columns set the form of the specification and the PDF of the match. Hover below for details.		
Power Meter Model		EPM			
Frequency		80 GHz			
Power		-26 dBm			2.51189E-06 W
Number of Readings Averaged		1			
Average mode		Normal			
Source of Uncertainty	Symbol	Value \pm	Probability Distribution	Divisor	Result
Mismatch Gain Between Generator and Sensor	M_u	$ \Gamma_{max} _g = 0.099$ $ \Gamma_{max} _s = 0.200$	C: Rayleigh C: Rayleigh	4.179	0.474%
Mismatch Gain Between Calibration Source and Sensor	M_{uc}	$ \Gamma _c = 0.029$ $ \Gamma_{max} _{sc} = 0.032$	A: Uniform inside Circle C: Rayleigh	2.431	0.038%
Power Meter Instrumentation Error	P_m	0.50%	Gaussian	2.000	0.250%
Power Meter Instrumentation Error During Calibration	P_{mc}	0.50%	Gaussian	2.000	0.250%
Power Meter Calibrator Output Power	P_{cal}	0.50%	Gaussian	2.000	0.250%
Zero Drift	D	4.000E-08	Gaussian	2.000	0.796%
Power Sensor Calibration Factor Uncertainties	K_b	5.300%	Gaussian	2.000	2.650%
Power Sensor Linearity	P_l	1.000%	Gaussian	2.000	0.500%
Zero Set	Z_s	2.000E-07	Gaussian	2.000	3.981%
Sensor Noise	N	4.500E-07	Gaussian	2.000	49.266%
Combined Uncertainty-RSSed =					49.51%
K =					2.00
* This term can be entered as the return loss (RL) in dB (Sxx), or VSWR, or as a reflection coefficient. e.g. RL = -15 dB is equivalent to VSWR = 1.43 is equivalent to $\rho = 0.178$ (reflection coefficient) Enter 1 for VSWR = 1, a perfect match, or 0 for RL = 0 dB, a lossless reflection					Expanded Uncertainty = 99.02%
					Upper Limit Uncertainty = + 2.989dB
					Lower Limit Uncertainty = - 20.092dB

7.2 OML WR12 Waveguide Offset Standards

+25 C	MIN	TYP	MAX
System Operating Frequency (WR-12) ¹	60 GHz	--	90 GHz
Length of Shim (1/4 wavelength)	.0533 +/- .0002 in		
Return Loss of Fixed Load, Adjustable Load	--	>35 dB	--
Damage Level	--	+13 dBm	--
Operating Temperature Range	+20 °C	+25 °C	+30 °C

¹Test Port Flange Configuration is compatible with MIL-DTL-3922/67D (UG387/U-M)